
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001579
8:50 11/12/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:05
PRESSURIZING 7:01
STERILIZING 8:40
136.0 C 220kPa 8:40
136.4 C 217kPa 9:10
136.2 C 220kPa 9:40
136.4 C 220kPa 10:10
136.4 C 222kPa 10:40
136.1 C 219kPa 11:10
136.5 C 222kPa 11:40
136.2 C 219kPa 12:10
VENTING 12:10
AIR DRYING 13:11
CYCLE COMPLETE 43:11

Digital Signature #
960DAE0311B20D86
