
Midmark M11 - v1.0.0
Total Cycles: 4595

Name: RITTER 1
Sterilizer ID: V2178198
05/30/2024 08:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.4 F 0.0
2:00 77.8 F 0.2
4:00 85.4 F 0.4
6:00 114.8 F 0.6
8:00 145.4 F 1.3
10:00 169.9 F 4.3
12:00 184.2 F 7.5
14:00 203.0 F 10.6
16:00 212.1 F 8.4
18:00 219.1 F 6.4
20:00 230.1 F 10.1
22:00 235.1 F 9.6
24:00 239.2 F 10.8
26:00 235.1 F 7.7
28:00 248.1 F 14.2
30:00 258.9 F 20.1
32:00 268.7 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.8 F 29.0
2:00 271.9 F 28.3
3:00 271.9 F 28.7
4:00 272.4 F 28.8

Min 271.4 F 28.1
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 209.0 F 0.0
10:00 207.9 F 0.0
15:00 206.5 F 0.0
20:00 208.7 F 0.0
25:00 212.4 F 0.0
30:00 213.9 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 33:17
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:11:23
CYCLE COMPLETE

05/30/2024 09:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____