
Midmark M11 - v1.0.5
Total Cycles: 3493

Name: RITTER2
Sterilizer ID: V2313228
04/03/2024 11:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 89.7 F 0.0
2:00 90.5 F 0.0
4:00 98.8 F 0.1
6:00 126.6 F 1.9
8:00 160.6 F 5.0
10:00 183.2 F 8.2
12:00 196.8 F 12.4
14:00 214.4 F 5.8
16:00 226.5 F 10.1
18:00 231.7 F 9.1
20:00 239.7 F 12.0
22:00 239.0 F 10.5
24:00 250.7 F 16.6
26:00 262.0 F 23.0
28:00 272.2 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.9 F 30.0
2:00 271.4 F 29.3
3:00 272.1 F 30.2
4:00 272.2 F 29.8

Min 271.1 F 29.2
Max 273.2 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.2 F 0.0
5:00 205.6 F 0.0
10:00 229.2 F 0.0
15:00 201.1 F 0.0
20:00 188.8 F 0.0
25:00 187.4 F 0.0
30:00 187.1 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 28:31
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:06:50
CYCLE COMPLETE

04/03/2024 12:40 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____