

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1348

Name: RITTER2  
Sterilizer ID: V2313228  
07/26/2022 12:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.7 F 0.0  
2:00 190.8 F 0.7  
4:00 208.4 F 6.0  
6:00 220.6 F 11.6  
8:00 225.9 F 7.2  
10:00 237.8 F 12.6  
12:00 238.3 F 10.8  
14:00 238.4 F 9.9  
16:00 252.1 F 16.9  
18:00 264.2 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.1  
1:00 272.8 F 29.6  
2:00 271.8 F 29.1  
3:00 272.9 F 29.7  
4:00 272.0 F 29.0  
  
Min 271.8 F 28.9  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.5 F 0.0  
5:00 207.8 F 0.0  
10:00 200.9 F 0.0  
15:00 207.1 F 0.0  
20:00 210.5 F 0.0  
25:00 213.6 F 0.0  
30:00 221.4 F 0.0  
Drying Complete

FILLING: 2:55  
HEATING: 19:56  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:04  
CYCLE COMPLETE

07/26/2022 01:16 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_