
Midmark M11 - v1.0.0
Total Cycles: 3522

Name: RITTER 1
Sterilizer ID: V2178198
07/13/2023 09:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.9 F 0.0
2:00 193.8 F 0.3
4:00 187.5 F 2.7
6:00 203.3 F 6.6
8:00 215.5 F 11.2
10:00 223.0 F 5.8
12:00 234.8 F 11.0
14:00 232.7 F 7.8
16:00 243.6 F 12.3
18:00 244.5 F 12.6
20:00 256.3 F 18.8
22:00 266.8 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.5 F 29.0
2:00 271.2 F 28.3
3:00 272.3 F 29.4
4:00 272.5 F 29.1

Min 271.0 F 28.3
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 216.4 F 0.0
10:00 230.4 F 0.0
15:00 229.2 F 0.0
20:00 230.2 F 0.0
25:00 225.6 F 0.0
30:00 219.2 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 23:35
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:02:18
CYCLE COMPLETE

07/13/2023 10:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____