

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2234

Name: RITTER 1  
Sterilizer ID: V2178198  
08/23/2022 09:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.1 F 0.0  
2:00 76.8 F 0.0  
4:00 90.8 F 0.0  
6:00 123.5 F 0.1  
8:00 152.7 F 2.2  
10:00 190.5 F 5.9  
12:00 204.6 F 10.2  
14:00 218.0 F 7.4  
16:00 226.6 F 7.5  
18:00 235.4 F 11.4  
20:00 232.5 F 7.5  
22:00 245.1 F 13.2  
24:00 243.0 F 11.5  
26:00 256.0 F 18.3  
28:00 267.3 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.0 F 28.5  
2:00 272.6 F 28.7  
3:00 272.3 F 28.4  
4:00 272.0 F 28.1

Min 271.5 F 27.9  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 234.0 F 0.0  
5:00 228.3 F 0.0  
Dry Cycle Aborted

FILLING: 0:47  
HEATING: 29:16  
STERILIZING: 4:00  
VENTING: 2:49  
DRYING: 7:23  
TOTAL CYCLE: 00:44:15

08/23/2022 10:29 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_