
Midmark M11 - v1.0.0
Total Cycles: 2028

Name: RITTER 1
Sterilizer ID: V2178198
06/16/2022 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.8 F 0.0
2:00 171.4 F 1.3
4:00 179.3 F 4.7
6:00 193.1 F 9.2
8:00 211.8 F 7.3
10:00 226.0 F 8.2
12:00 237.2 F 13.3
14:00 236.7 F 10.2
16:00 236.1 F 8.1
18:00 248.2 F 14.6
20:00 259.8 F 21.1
22:00 270.5 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.3 F 28.9
2:00 271.9 F 29.2
3:00 271.7 F 28.5
4:00 272.5 F 29.3

Min 271.3 F 28.5
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 214.6 F 0.0
10:00 220.1 F 0.0
15:00 221.9 F 0.0
20:00 222.2 F 0.0
25:00 223.5 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:48
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:01:02
CYCLE COMPLETE

06/16/2022 10:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____