
Midmark M11 - v1.0.0
Total Cycles: 2442

Name: RITTER 1
Sterilizer ID: V2178198
10/17/2022 01:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.2 F 0.0
2:00 211.2 F 4.8
4:00 219.9 F 10.4
6:00 224.8 F 5.8
8:00 236.7 F 11.7
10:00 236.0 F 9.3
12:00 238.0 F 9.2
14:00 248.7 F 14.8
16:00 260.8 F 21.6
18:00 271.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.6
1:00 272.8 F 29.2
2:00 271.7 F 28.5
3:00 272.8 F 29.3
4:00 272.1 F 28.7

Min 271.7 F 28.4
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.3 F 0.0
5:00 233.6 F 0.0
10:00 227.2 F 0.0
15:00 228.6 F 0.0
20:00 230.1 F 0.0
25:00 228.1 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 2:45
HEATING: 18:35
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

10/17/2022 02:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____