
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000136
17:52 08/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:27
PRESSURIZING 3:31
STERILIZING 4:59
136.0 C 222kPa 4:59
136.3 C 225kPa 5:29
136.3 C 225kPa 5:59
136.2 C 224kPa 6:29
136.4 C 225kPa 6:59
136.3 C 223kPa 7:29
136.2 C 224kPa 7:59
136.4 C 226kPa 8:29
136.4 C 226kPa 8:29
VENTING 8:29
AIR DRYING 9:33
CYCLE COMPLETE 15:48

Digital Signature #
C6CB8035E7EC020B
