
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005793
8:12 20/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:57
STERILIZING 3:06
136.0 C 220kPa 3:06
136.1 C 220kPa 3:36
136.2 C 221kPa 4:06
136.3 C 220kPa 4:36
136.2 C 223kPa 5:06
136.2 C 223kPa 5:36
136.0 C 221kPa 6:06
136.4 C 222kPa 6:36
136.4 C 222kPa 6:36
VENTING 6:36
AIR DRYING 7:30
CYCLE COMPLETE 7:56

Digital Signature #
1D617B569D999438
