
Midmark M11 - v1.0.0
Total Cycles: 1656

Name: RITTER 1
Sterilizer ID: V2178198
02/22/2022 01:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.5 F 0.0
2:00 122.3 F 0.3
4:00 144.9 F 2.1
6:00 172.8 F 4.8
8:00 192.9 F 8.7
10:00 212.0 F 13.5
12:00 223.2 F 7.0
14:00 235.6 F 11.8
16:00 234.7 F 8.7
18:00 240.4 F 10.2
20:00 246.3 F 13.3
22:00 258.5 F 20.1
24:00 269.9 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.3
1:00 272.8 F 29.1
2:00 271.6 F 28.3
3:00 272.4 F 28.8
4:00 272.0 F 28.7

Min 271.5 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 211.8 F 0.0
10:00 229.3 F 0.0
15:00 226.3 F 0.0
20:00 232.3 F 0.0
25:00 226.7 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 24:45
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:03:05
CYCLE COMPLETE

02/22/2022 02:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____