
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001508
9:42 18/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:42
STERILIZING 3:42
136.4 C 224kPa 3:42
136.4 C 226kPa 4:12
136.1 C 223kPa 4:42
136.3 C 224kPa 5:12
136.4 C 226kPa 5:42
136.2 C 222kPa 6:12
136.4 C 225kPa 6:42
136.2 C 223kPa 7:12
136.2 C 223kPa 7:12
VENTING 7:12
AIR DRYING 8:03
CYCLE COMPLETE 8:10

Digital Signature #
B4096FBB08D252E7
