

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 001100  
16:34 26/05/2022  
WRAPPED  
135 C for 10 min.  
CYCLE START 0:00  
CONDITIONING 3:42  
PRESSURIZING 6:09  
STERILIZING 7:22  
136.4 C 225kPa 7:22  
136.3 C 224kPa 7:52  
136.4 C 225kPa 8:22  
136.4 C 224kPa 8:52  
136.3 C 225kPa 9:22  
136.4 C 225kPa 9:52  
136.4 C 224kPa 10:22  
136.4 C 225kPa 10:52  
136.2 C 223kPa 11:22  
136.5 C 226kPa 11:52  
136.2 C 223kPa 12:22  
136.3 C 224kPa 12:52  
136.4 C 225kPa 13:22  
136.4 C 226kPa 13:52  
136.4 C 225kPa 14:22  
136.2 C 225kPa 14:52  
136.4 C 226kPa 15:22  
136.3 C 225kPa 15:52  
136.2 C 224kPa 16:22  
136.2 C 224kPa 16:52  
136.4 C 226kPa 17:22  
136.4 C 226kPa 17:22  
VENTING 17:22  
AIR DRYING 18:20  
CYCLE COMPLETE 18:31  
-----

Digital Signature #  
26A61DB9C25E30E8  
-----