

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.0uS / 0.6ppm  
CYCLE NUMBER 001511  
13:54 10/18/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:32  
PRESSURIZING 6:14  
STERILIZING 7:43  
136.0 C 220kPa 7:43  
136.5 C 223kPa 8:13  
136.5 C 220kPa 8:43  
136.4 C 223kPa 9:13  
136.4 C 223kPa 9:43  
136.3 C 222kPa 10:13  
136.5 C 224kPa 10:43  
136.5 C 224kPa 11:13  
VENTING 11:13  
AIR DRYING 12:19  
CYCLE COMPLETE 42:19

-----  
Digital Signature #  
15F9486B909B72CC  
-----