
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 002697
10:35 17/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:46
STERILIZING 3:48
136.4 C 223kPa 3:48
136.2 C 219kPa 4:18
136.4 C 224kPa 4:48
136.3 C 220kPa 5:18
136.1 C 222kPa 5:48
136.3 C 223kPa 6:18
136.4 C 224kPa 6:48
136.3 C 222kPa 7:18
136.3 C 222kPa 7:18
VENTING 7:18
AIR DRYING 8:08
CYCLE COMPLETE 8:10

Digital Signature #
5F351F22F11ED9A4
