
Midmark M11 - v1.0.5
Total Cycles: 2385

Name: RITTER2
Sterilizer ID: V2313228
05/17/2023 12:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.5 F 0.0
2:00 184.6 F 3.9
4:00 199.7 F 8.9
6:00 222.2 F 7.8
8:00 232.5 F 10.0
10:00 234.0 F 8.2
12:00 241.5 F 11.4
14:00 246.6 F 13.8
16:00 259.3 F 20.9
18:00 270.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.7 F 29.7
2:00 272.0 F 29.1
3:00 271.7 F 29.3
4:00 272.7 F 29.7

Min 271.4 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.8 F 0.0
5:00 218.8 F 0.0
10:00 228.3 F 0.0
15:00 233.0 F 0.0
20:00 223.5 F 0.0
25:00 221.4 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 18:49
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:08
CYCLE COMPLETE

05/17/2023 01:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____