
Midmark M11 - v1.0.5
Total Cycles: 4051

Name: RITTER2
Sterilizer ID: V2313228
10/24/2024 09:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 70.3 F 0.0
4:00 76.3 F 0.0
6:00 98.5 F 0.0
8:00 148.8 F 0.1
10:00 181.0 F 0.4
12:00 202.8 F 2.4
14:00 212.3 F 5.5
16:00 222.0 F 8.9
18:00 232.5 F 12.7
20:00 232.5 F 8.7
22:00 241.7 F 12.0
24:00 242.7 F 12.1
26:00 241.4 F 11.1
28:00 254.1 F 17.6
30:00 265.3 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.7 F 29.2
2:00 271.5 F 28.6
3:00 272.7 F 29.2
4:00 271.6 F 28.5

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.0 F 0.0
5:00 218.6 F 0.0
10:00 190.3 F 0.0
15:00 187.0 F 0.0
20:00 184.8 F 0.0
25:00 184.8 F 0.0
30:00 198.6 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:50
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:10:03
CYCLE COMPLETE

10/24/2024 10:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____