
Midmark M11 - v1.0.5
Total Cycles: 4322

Name: RITTER2
Sterilizer ID: V2313228
01/31/2025 10:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 73.8 F 0.0
4:00 81.4 F 0.4
6:00 101.9 F 2.3
8:00 144.4 F 5.2
10:00 172.6 F 8.6
12:00 183.0 F 12.2
14:00 203.9 F 3.9
16:00 214.9 F 7.3
18:00 226.5 F 11.0
20:00 226.1 F 7.0
22:00 239.1 F 11.8
24:00 237.3 F 9.8
26:00 249.8 F 15.8
28:00 260.5 F 21.9
30:00 270.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.8 F 29.8
2:00 271.5 F 29.2
3:00 272.8 F 29.8
4:00 271.5 F 29.1

Min 271.4 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.4 F 0.0
5:00 197.2 F 0.0
10:00 206.5 F 0.0
15:00 209.1 F 0.0
20:00 217.3 F 0.0
25:00 220.0 F 0.0
30:00 220.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:50
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:09:11
CYCLE COMPLETE

01/31/2025 11:10 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____