
Midmark M11 - v1.0.0
Total Cycles: 747

Name: RITTER 1
Sterilizer ID: V2178198
03/25/2021 10:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.7 F 0.0
2:00 170.3 F 1.6
4:00 178.4 F 5.2
6:00 197.0 F 10.3
8:00 221.9 F 5.0
10:00 224.4 F 10.5
12:00 230.9 F 7.0
14:00 242.2 F 12.6
16:00 243.2 F 12.0
18:00 255.0 F 18.7
20:00 266.6 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.3
1:00 272.6 F 29.6
2:00 272.4 F 30.0
3:00 271.6 F 29.2
4:00 273.0 F 29.9

Min 271.4 F 29.2
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 229.5 F 0.0
10:00 225.8 F 0.0
15:00 224.0 F 0.0
20:00 219.1 F 0.0
25:00 217.3 F 0.0
30:00 213.8 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

03/25/2021 11:01 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____