
Midmark M11 - v1.0.5
Total Cycles: 4279

Name: RITTER2
Sterilizer ID: V2313228
01/22/2025 12:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.9 F 0.0
2:00 163.5 F 2.0
4:00 163.9 F 5.6
6:00 189.5 F 9.9
8:00 204.1 F 6.4
10:00 216.0 F 8.1
12:00 229.2 F 12.7
14:00 234.3 F 9.7
16:00 238.6 F 9.6
18:00 246.3 F 14.1
20:00 257.9 F 20.3
22:00 268.1 F 26.9
24:00 272.0 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.7
1:00 272.4 F 29.6
2:00 272.7 F 29.8
3:00 272.8 F 30.1
4:00 272.4 F 30.3

Min 272.0 F 29.6
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 180.5 F 0.0
10:00 184.6 F 0.0
15:00 184.2 F 0.0
20:00 181.1 F 0.0
25:00 180.8 F 0.0
30:00 178.3 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 24:01
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:03:06
CYCLE COMPLETE

01/22/2025 01:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____