
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 003665
8:02 18/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:10
STERILIZING 4:19
136.4 C 224kPa 4:19
136.4 C 224kPa 4:49
136.3 C 220kPa 5:19
136.2 C 217kPa 5:49
136.3 C 222kPa 6:19
136.4 C 223kPa 6:49
136.2 C 223kPa 7:19
136.4 C 223kPa 7:49
136.4 C 223kPa 7:49
VENTING 7:49
AIR DRYING 8:37
CYCLE COMPLETE 9:00

Digital Signature #
FC5465809A78F087
