
Midmark M11 - v1.0.0
Total Cycles: 4184

Name: RITTER 1
Sterilizer ID: V2178198
01/09/2024 03:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.9 F 0.0
2:00 69.5 F 0.3
4:00 77.7 F 1.1
6:00 100.5 F 2.7
8:00 128.6 F 4.8
10:00 154.3 F 7.4
12:00 176.4 F 10.4
14:00 197.2 F 10.8
16:00 204.4 F 4.3
18:00 213.7 F 7.4
20:00 223.9 F 10.9
22:00 229.1 F 7.3
24:00 237.7 F 10.7
26:00 235.6 F 8.5
28:00 245.9 F 13.5
30:00 256.0 F 18.9
32:00 265.6 F 24.7
34:00 271.2 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.1 F 29.2
2:00 272.5 F 29.5
3:00 272.2 F 29.0
4:00 271.5 F 28.6

Min 271.1 F 28.5
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.8 F 0.0
5:00 230.8 F 0.0
10:00 209.7 F 0.0
15:00 218.9 F 0.0
20:00 226.7 F 0.0
25:00 230.5 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 34:02
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:12:38
CYCLE COMPLETE

01/09/2024 05:11 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____