

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1083

Name: RITTER 1  
Sterilizer ID: V2178198  
08/10/2021 06:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.9 F 0.0  
2:00 79.5 F 0.0  
4:00 83.7 F 0.0  
6:00 117.4 F 0.1  
8:00 151.8 F 0.1  
10:00 187.8 F 0.4  
12:00 203.0 F 4.1  
14:00 218.1 F 7.5  
16:00 225.2 F 11.1  
18:00 229.2 F 6.2  
20:00 240.5 F 11.2  
22:00 235.6 F 8.3  
24:00 241.1 F 10.2  
26:00 247.5 F 13.7  
28:00 260.0 F 20.5  
30:00 271.1 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.0  
1:00 272.4 F 28.3  
2:00 272.4 F 28.7  
3:00 271.6 F 27.9  
4:00 272.7 F 28.6

Min 271.4 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.9 F 0.0  
5:00 228.7 F 0.0  
10:00 236.4 F 0.0  
15:00 216.7 F 0.0  
20:00 182.8 F 0.0  
25:00 172.9 F 0.0  
30:00 211.2 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 30:40  
STERILIZING: 4:00  
VENTING: 2:54  
DRYING: 30:00  
TOTAL CYCLE: 01:08:25  
CYCLE COMPLETE

08/10/2021 07:54 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01