
Midmark M11 - v1.0.0
Total Cycles: 630

Name: MIDMARK
Sterilizer ID: V2178198
01/26/2021 12:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.4 F 0.0
2:00 173.0 F 2.4
4:00 175.6 F 6.1
6:00 236.7 F 11.8
8:00 226.2 F 7.9
10:00 237.2 F 11.2
12:00 242.3 F 11.5
14:00 239.1 F 9.7
16:00 253.5 F 16.8
18:00 265.6 F 24.1
20:00 272.9 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 273.3 F 29.5
2:00 271.4 F 28.6
3:00 272.9 F 29.6
4:00 271.7 F 29.1

Min 271.1 F 28.6
Max 273.4 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.1 F 0.0
5:00 234.3 F 0.0
10:00 222.4 F 0.0
15:00 225.3 F 0.0
20:00 228.8 F 0.0
25:00 229.3 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 20:25
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

01/26/2021 01:00 PM
ACCEPTED: X
REJECTED:
Operator: U01