
Midmark M11 - v1.0.0
Total Cycles: 4897

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2024 08:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.2 F 0.0
2:00 98.7 F 0.0
4:00 91.3 F 0.0
6:00 121.7 F 0.0
8:00 148.1 F 0.3
10:00 176.1 F 0.6
12:00 197.9 F 2.9
14:00 219.0 F 6.1
16:00 229.5 F 9.6
18:00 240.7 F 13.5
20:00 235.0 F 8.7
22:00 245.3 F 13.0
24:00 242.2 F 11.3
26:00 239.0 F 9.6
28:00 252.9 F 16.4
30:00 264.5 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.7 F 28.5
2:00 271.6 F 27.7
3:00 272.7 F 28.7
4:00 272.0 F 28.1

Min 271.4 F 27.7
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.5 F 0.0
5:00 218.5 F 0.0
10:00 188.8 F 0.0
15:00 197.1 F 0.0
20:00 205.9 F 0.0
25:00 233.9 F 0.0
30:00 233.6 F 0.0
Drying Complete

FILLING: 0:40
HEATING: 31:57
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:09:59
CYCLE COMPLETE

09/12/2024 09:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____