

-----  
Midmark M11 - v1.0.0  
Total Cycles: 825

Name: RITTER 1  
Sterilizer ID: V2178198  
05/19/2021 03:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.1 F 0.0  
2:00 191.9 F 3.9  
4:00 206.1 F 9.3  
6:00 223.3 F 6.1  
8:00 235.9 F 11.1  
10:00 235.4 F 8.7  
12:00 238.7 F 9.2  
14:00 249.7 F 14.9  
16:00 263.1 F 22.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.1 F 28.2  
2:00 272.5 F 28.7  
3:00 272.3 F 28.9  
4:00 271.7 F 28.1  
  
Min 271.6 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.1 F 0.0  
5:00 238.8 F 0.0  
10:00 229.3 F 0.0  
15:00 224.8 F 0.0  
20:00 207.5 F 0.0  
25:00 201.4 F 0.0  
Dry Cycle Aborted

FILLING: 2:15  
HEATING: 17:55  
STERILIZING: 4:00  
VENTING: 2:57  
DRYING: 26:14  
TOTAL CYCLE: 00:53:21

05/19/2021 04:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_