
Midmark M11 - v1.0.0
Total Cycles: 4590

Name: RITTER 1
Sterilizer ID: V2178198
05/28/2024 01:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.3 F 0.0
2:00 107.0 F 1.0
4:00 126.6 F 3.1
6:00 151.5 F 5.8
8:00 174.1 F 8.9
10:00 196.1 F 11.9
12:00 208.5 F 3.8
14:00 215.4 F 7.3
16:00 224.9 F 11.0
18:00 228.2 F 7.0
20:00 237.9 F 11.0
22:00 234.0 F 8.2
24:00 247.0 F 14.1
26:00 257.4 F 19.6
28:00 267.0 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.4 F 29.5
2:00 272.3 F 29.1
3:00 271.4 F 28.6
4:00 271.4 F 28.9

Min 271.2 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.0 F 0.0
5:00 203.5 F 0.0
10:00 211.0 F 0.0
15:00 209.2 F 0.0
20:00 219.5 F 0.0
25:00 230.1 F 0.0
30:00 233.5 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 29:43
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:08:34
CYCLE COMPLETE

05/28/2024 02:48 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____