

-----  
Midmark M11 - v1.0.5  
Total Cycles: 224

Name: RITTER2  
Sterilizer ID: V2313228  
07/19/2021 01:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 108.2 F 0.0  
2:00 109.5 F 0.8  
4:00 115.7 F 2.8  
6:00 140.9 F 5.7  
8:00 168.2 F 9.2  
10:00 189.6 F 13.0  
12:00 207.5 F 5.6  
14:00 221.9 F 9.7  
16:00 234.0 F 10.6  
18:00 235.4 F 11.3  
20:00 235.5 F 9.5  
22:00 247.5 F 15.2  
24:00 258.0 F 21.0  
26:00 267.7 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 30.0  
1:00 272.7 F 30.5  
2:00 271.4 F 29.9  
3:00 272.5 F 30.7  
4:00 272.1 F 30.2

Min 271.3 F 29.8  
Max 273.2 F 30.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.9 F 0.0  
5:00 220.1 F 0.0  
10:00 228.5 F 0.0  
15:00 223.0 F 0.0  
20:00 200.3 F 0.0  
25:00 186.3 F 0.0  
30:00 181.6 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 27:28  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:05:55  
CYCLE COMPLETE

07/19/2021 02:36 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01