
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000788
15:41 10/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:37
PRESSURIZING 6:17
STERILIZING 7:39
136.0 C 220kPa 7:39
136.1 C 221kPa 8:09
135.9 C 219kPa 8:39
136.3 C 223kPa 9:09
136.5 C 221kPa 9:39
136.4 C 223kPa 10:09
136.4 C 223kPa 10:39
136.2 C 222kPa 11:09
136.2 C 222kPa 11:09
VENTING 11:09
AIR DRYING 12:11
CYCLE COMPLETE 25:20

Digital Signature #
F68EF934D0E11315
