
Midmark M11 - v1.0.5
Total Cycles: 2300

Name: RITTER2
Sterilizer ID: V2313228
04/25/2023 12:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.2 F 0.0
2:00 148.6 F 2.7
4:00 167.4 F 6.2
6:00 189.8 F 11.2
8:00 213.2 F 5.5
10:00 228.3 F 10.7
12:00 229.2 F 7.7
14:00 240.3 F 11.7
16:00 244.2 F 13.1
18:00 257.0 F 19.9
20:00 268.6 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.9 F 30.1
2:00 271.8 F 29.4
3:00 272.3 F 30.1
4:00 272.9 F 29.9

Min 271.3 F 29.3
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.1 F 0.0
5:00 209.8 F 0.0
10:00 216.7 F 0.0
15:00 228.4 F 0.0
20:00 227.2 F 0.0
25:00 229.0 F 0.0
30:00 230.5 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 21:12
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

04/25/2023 01:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____