
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 001475
12:42 10/06/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:20
PRESSURIZING 7:15
STERILIZING 8:57
136.0 C 220kPa 8:57
136.2 C 220kPa 9:27
136.3 C 221kPa 9:57
136.4 C 222kPa 10:27
136.4 C 223kPa 10:58
136.1 C 220kPa 11:28
136.0 C 220kPa 11:58
136.4 C 223kPa 12:27
VENTING 12:27
AIR DRYING 13:30
CYCLE COMPLETE 24:26

Digital Signature #
6CC7B90B5F593EAA
