
Midmark M11 - v1.0.5
Total Cycles: 3582

Name: RITTER2
Sterilizer ID: V2313228
05/02/2024 11:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.7 F 0.0
2:00 144.2 F 1.9
4:00 151.3 F 5.0
6:00 181.8 F 9.1
8:00 199.7 F 9.9
10:00 219.2 F 7.4
12:00 230.9 F 12.1
14:00 231.6 F 8.8
16:00 239.3 F 10.9
18:00 244.1 F 13.2
20:00 256.1 F 19.5
22:00 266.9 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.4
1:00 272.6 F 30.0
2:00 271.4 F 29.3
3:00 272.0 F 30.1
4:00 272.5 F 29.8

Min 271.1 F 29.2
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.6 F 0.0
5:00 214.3 F 0.0
10:00 203.4 F 0.0
15:00 200.7 F 0.0
20:00 204.8 F 0.0
25:00 206.9 F 0.0
30:00 204.6 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 23:33
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:02:11
CYCLE COMPLETE

05/02/2024 12:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____