

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4438

Name: RITTER2  
Sterilizer ID: V2313228  
03/20/2025 12:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 83.0 F 0.0  
2:00 83.4 F 0.4  
4:00 93.1 F 1.7  
6:00 116.6 F 4.3  
8:00 149.2 F 7.5  
10:00 171.7 F 11.2  
12:00 203.8 F 3.9  
14:00 217.0 F 6.9  
16:00 226.8 F 10.6  
18:00 228.4 F 8.3  
20:00 237.2 F 10.9  
22:00 234.0 F 8.2  
24:00 246.7 F 14.3  
26:00 257.4 F 20.1  
28:00 267.0 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.6  
1:00 271.4 F 29.5  
2:00 271.5 F 29.7  
3:00 271.5 F 29.9  
4:00 271.5 F 30.0

Min 271.3 F 29.5  
Max 272.1 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.9 F 0.0  
5:00 223.5 F 0.0  
10:00 221.8 F 0.0  
15:00 218.3 F 0.0  
20:00 217.7 F 0.0  
25:00 227.4 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 29:20  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:07:57  
CYCLE COMPLETE

03/20/2025 01:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_