

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2525

Name: RITTER 1  
Sterilizer ID: V2178198  
11/07/2022 08:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.7 F 0.0  
2:00 179.6 F 4.6  
4:00 196.8 F 9.6  
6:00 218.4 F 5.9  
8:00 228.0 F 9.3  
10:00 234.9 F 9.0  
12:00 242.9 F 12.5  
14:00 241.9 F 11.2  
16:00 255.4 F 18.2  
18:00 267.1 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.5 F 28.9  
2:00 271.5 F 28.4  
3:00 272.5 F 29.0  
4:00 271.6 F 28.3  
  
Min 271.3 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.6 F 0.0  
5:00 232.7 F 0.0  
10:00 226.8 F 0.0  
15:00 227.9 F 0.0  
20:00 227.6 F 0.0  
25:00 230.3 F 0.0  
30:00 228.5 F 0.0  
  
Drying Complete

FILLING: 1:40  
HEATING: 19:23  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:13  
CYCLE COMPLETE

11/07/2022 09:13 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_