
Midmark M11 - v1.0.0
Total Cycles: 1888

Name: RITTER 1
Sterilizer ID: V2178198
04/28/2022 03:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 216.5 F 0.0
2:00 209.3 F 3.8
4:00 214.9 F 9.3
6:00 228.6 F 6.8
8:00 236.0 F 10.6
10:00 236.1 F 7.6
12:00 242.6 F 11.2
14:00 246.9 F 13.3
16:00 259.3 F 20.2
18:00 270.5 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.4 F 28.7
2:00 271.3 F 28.1
3:00 272.4 F 28.7
4:00 271.4 F 28.0

Min 271.2 F 27.9
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 232.3 F 0.0
10:00 231.8 F 0.0
15:00 227.6 F 0.0
20:00 227.5 F 0.0
25:00 229.2 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 2:07
HEATING: 18:49
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

04/28/2022 04:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____