
Midmark M11 - v1.0.0
Total Cycles: 3618

Name: RITTER 1
Sterilizer ID: V2178198
08/04/2023 01:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.8 F 0.0
2:00 96.8 F 0.0
4:00 109.5 F 0.1
6:00 144.6 F 1.5
8:00 172.5 F 4.3
10:00 199.3 F 7.2
12:00 207.8 F 10.5
14:00 223.8 F 8.5
16:00 226.7 F 7.4
18:00 236.5 F 11.5
20:00 233.0 F 7.5
22:00 243.3 F 12.8
24:00 241.3 F 10.9
26:00 252.8 F 16.6
28:00 262.8 F 22.5
30:00 272.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.6 F 29.0
2:00 271.8 F 28.3
3:00 271.6 F 28.6
4:00 272.5 F 29.0

Min 271.2 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 192.3 F 0.0
10:00 218.6 F 0.0
15:00 226.1 F 0.0
20:00 227.9 F 0.0
25:00 225.6 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:28
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:08:53
CYCLE COMPLETE

08/04/2023 02:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____