

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4978

Name: RITTER 1  
Sterilizer ID: V2178198  
10/08/2024 04:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.4 F 0.0  
2:00 119.9 F 0.2  
4:00 146.6 F 0.8  
6:00 173.5 F 3.3  
8:00 194.4 F 7.2  
10:00 214.2 F 11.2  
12:00 218.1 F 5.0  
14:00 228.2 F 9.8  
16:00 234.2 F 9.0  
18:00 240.5 F 11.5  
20:00 237.6 F 9.5  
22:00 251.3 F 15.9  
24:00 262.6 F 22.4  
26:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.6 F 29.1  
2:00 271.8 F 28.4  
3:00 271.8 F 28.7  
4:00 272.6 F 29.1

Min 271.3 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.4 F 0.0  
5:00 203.4 F 0.0  
10:00 191.5 F 0.0  
15:00 191.8 F 0.0  
20:00 218.7 F 0.0  
25:00 233.0 F 0.0  
30:00 229.7 F 0.0

Drying Complete

FILLING: 0:50  
HEATING: 26:24  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:04:45  
CYCLE COMPLETE

10/08/2024 05:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_