
Midmark M11 - v1.0.5
Total Cycles: 3767

Name: RITTER2
Sterilizer ID: V2313228
07/11/2024 08:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 70.9 F 0.0
4:00 77.3 F 0.0
6:00 105.1 F 0.2
8:00 148.7 F 2.2
10:00 173.7 F 5.1
12:00 190.1 F 8.8
14:00 207.9 F 12.8
16:00 217.5 F 5.3
18:00 228.0 F 9.0
20:00 237.8 F 13.2
22:00 237.0 F 10.1
24:00 237.3 F 9.2
26:00 247.4 F 14.3
28:00 258.6 F 20.5
30:00 269.1 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.3
2:00 271.9 F 29.2
3:00 272.4 F 29.1
4:00 272.0 F 29.4

Min 271.5 F 28.6
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 225.6 F 0.0
10:00 198.8 F 0.0
15:00 192.5 F 0.0
20:00 190.9 F 0.0
25:00 192.7 F 0.0
30:00 198.5 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:06
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:09:18
CYCLE COMPLETE

07/11/2024 09:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____