
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007407
7:09 22/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:35
PRESSURIZING 3:38
STERILIZING 4:44
136.4 C 222kPa 4:44
136.4 C 224kPa 5:14
136.5 C 223kPa 5:44
135.9 C 219kPa 6:14
136.2 C 223kPa 6:44
135.9 C 219kPa 7:14
136.2 C 222kPa 7:44
136.2 C 217kPa 8:14
136.2 C 217kPa 8:14
VENTING 8:14
AIR DRYING 9:05
CYCLE COMPLETE 30:40

Digital Signature #
EE82B011EB6844B4
