
Midmark M11 - v1.0.0
Total Cycles: 4492

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2024 11:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.0 F 0.0
2:00 110.2 F 1.0
4:00 127.2 F 3.2
6:00 148.8 F 6.0
8:00 167.5 F 9.3
10:00 191.5 F 13.1
12:00 209.3 F 5.4
14:00 224.2 F 9.2
16:00 232.5 F 13.4
18:00 233.3 F 9.1
20:00 240.6 F 11.2
22:00 242.9 F 11.9
24:00 254.1 F 17.5
26:00 263.9 F 23.5
28:00 271.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.5 F 29.1
2:00 272.4 F 29.5
3:00 271.9 F 28.8
4:00 272.5 F 29.4

Min 271.7 F 28.8
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 217.5 F 0.0
10:00 208.4 F 0.0
15:00 211.1 F 0.0
20:00 228.9 F 0.0
25:00 232.6 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 28:55
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:07:32
CYCLE COMPLETE

04/25/2024 12:33 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____