

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2714

Name: RITTER 1  
Sterilizer ID: V2178198  
12/21/2022 05:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.6 F 0.0  
2:00 200.0 F 4.1  
4:00 211.4 F 9.5  
6:00 220.0 F 6.2  
8:00 235.0 F 11.0  
10:00 235.6 F 8.4  
12:00 240.0 F 9.8  
14:00 249.6 F 14.5  
16:00 262.9 F 22.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.8  
1:00 272.1 F 28.0  
2:00 272.7 F 28.5  
3:00 271.5 F 27.6  
4:00 272.7 F 28.4  
  
Min 271.4 F 27.6  
Max 272.8 F 28.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.1 F 0.0  
5:00 212.1 F 0.0  
10:00 226.7 F 0.0  
15:00 229.3 F 0.0  
20:00 224.3 F 0.0  
25:00 220.2 F 0.0  
30:00 217.1 F 0.0  
Drying Complete

FILLING: 1:54  
HEATING: 17:60  
STERILIZING: 4:00  
VENTING: 2:60  
DRYING: 30:00  
TOTAL CYCLE: 00:56:53  
CYCLE COMPLETE

12/21/2022 06:41 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_