
Midmark M11 - v1.0.5
Total Cycles: 4207

Name: RITTER2
Sterilizer ID: V2313228
12/12/2024 11:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.9 F 0.0
2:00 144.2 F 1.5
4:00 155.5 F 4.1
6:00 176.0 F 8.4
8:00 197.7 F 12.9
10:00 213.3 F 6.8
12:00 227.4 F 11.6
14:00 229.7 F 8.0
16:00 242.1 F 13.0
18:00 242.5 F 12.2
20:00 254.7 F 18.6
22:00 265.8 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.7 F 30.1
2:00 271.9 F 29.4
3:00 271.4 F 29.5
4:00 272.7 F 30.1

Min 271.2 F 29.1
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 206.6 F 0.0
10:00 200.8 F 0.0
15:00 198.9 F 0.0
20:00 234.2 F 0.0
25:00 229.5 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 23:47
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:02:12
CYCLE COMPLETE

12/12/2024 12:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____