
Midmark M11 - v1.0.5
Total Cycles: 394

Name: RITTER2
Sterilizer ID: V2313228
09/27/2021 08:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.5 F 0.0
2:00 119.8 F 0.7
4:00 126.1 F 3.1
6:00 158.6 F 6.5
8:00 184.8 F 10.7
10:00 201.0 F 5.9
12:00 219.1 F 9.1
14:00 228.2 F 9.9
16:00 236.4 F 11.2
18:00 236.0 F 9.4
20:00 249.3 F 15.8
22:00 260.8 F 22.4
24:00 271.5 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 272.7 F 30.1
2:00 271.6 F 29.6
3:00 272.7 F 30.2
4:00 271.6 F 29.4

Min 271.5 F 29.4
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.0 F 0.0
5:00 221.2 F 0.0
10:00 222.0 F 0.0
15:00 224.4 F 0.0
20:00 224.6 F 0.0
25:00 225.9 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:37
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:02:38
CYCLE COMPLETE

09/27/2021 09:29 AM
ACCEPTED: X
REJECTED:
Operator: U01