
Midmark M11 - v1.0.5
Total Cycles: 767

Name: RITTER2
Sterilizer ID: V2313228
02/08/2022 07:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.7 F 0.0
2:00 77.3 F 0.0
4:00 82.0 F 0.0
6:00 111.7 F 0.1
8:00 148.1 F 0.3
10:00 177.6 F 1.7
12:00 193.7 F 4.9
14:00 206.3 F 8.5
16:00 218.0 F 12.5
18:00 226.4 F 7.4
20:00 236.7 F 12.0
22:00 235.7 F 9.4
24:00 239.4 F 10.1
26:00 247.4 F 14.2
28:00 258.8 F 20.5
30:00 269.2 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.3 F 29.2
2:00 271.7 F 29.1
3:00 272.1 F 29.0
4:00 272.0 F 29.3

Min 271.4 F 28.6
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.7 F 0.0
5:00 225.6 F 0.0
10:00 204.0 F 0.0
15:00 195.2 F 0.0
20:00 191.5 F 0.0
25:00 215.2 F 0.0
30:00 212.7 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:06
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:08
CYCLE COMPLETE

02/08/2022 08:49 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____