
Midmark M11 - v1.0.0
Total Cycles: 4932

Name: RITTER 1
Sterilizer ID: V2178198
09/23/2024 03:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.4 F 0.0
2:00 149.0 F 0.1
4:00 169.0 F 0.5
6:00 195.3 F 3.7
8:00 209.6 F 7.9
10:00 223.1 F 12.2
12:00 226.6 F 7.1
14:00 239.2 F 12.5
16:00 238.1 F 10.2
18:00 235.9 F 8.6
20:00 250.6 F 15.6
22:00 262.4 F 22.3
24:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.9
2:00 271.8 F 28.2
3:00 272.1 F 28.8
4:00 272.5 F 28.8

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.6 F 0.0
5:00 203.7 F 0.0
10:00 199.7 F 0.0
15:00 204.9 F 0.0
20:00 214.6 F 0.0
25:00 224.8 F 0.0
30:00 232.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:20
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:02:47
CYCLE COMPLETE

09/23/2024 04:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____