

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1526

Name: RITTER 1  
Sterilizer ID: V2178198  
01/13/2022 03:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.1 F 0.0  
2:00 153.2 F 2.9  
4:00 178.4 F 6.7  
6:00 202.1 F 11.4  
8:00 218.8 F 4.6  
10:00 227.0 F 9.6  
12:00 234.7 F 9.4  
14:00 241.3 F 11.6  
16:00 238.7 F 9.7  
18:00 252.3 F 16.4  
20:00 263.8 F 23.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.2 F 28.5  
2:00 272.5 F 29.0  
3:00 271.6 F 28.5  
4:00 272.1 F 28.7  
  
Min 271.5 F 28.3  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.6 F 0.0  
5:00 228.7 F 0.0  
10:00 223.0 F 0.0  
15:00 200.6 F 0.0  
Dry Cycle Aborted

FILLING: 1:27  
HEATING: 21:59  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 19:50  
TOTAL CYCLE: 00:50:27

01/13/2022 04:15 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02