
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 001558
11:12 11/04/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:22
PRESSURIZING 5:26
STERILIZING 7:01
136.0 C 220kPa 7:01
136.1 C 220kPa 7:31
136.5 C 223kPa 8:01
136.4 C 223kPa 8:31
136.1 C 221kPa 9:01
136.4 C 223kPa 9:31
136.2 C 221kPa 10:01
136.3 C 222kPa 10:31
VENTING 10:31
AIR DRYING 11:31
CYCLE COMPLETE 41:31

Digital Signature #
5D4AEED8D2719461
