
Midmark M11 - v1.0.0
Total Cycles: 4279

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2024 12:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.3 F 0.0
2:00 164.9 F 2.6
4:00 174.9 F 6.2
6:00 198.0 F 10.4
8:00 219.5 F 4.7
10:00 223.3 F 8.8
12:00 238.3 F 12.6
14:00 238.0 F 10.5
16:00 235.8 F 8.2
18:00 249.4 F 15.0
20:00 260.9 F 21.4
22:00 271.5 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.1 F 29.1
2:00 272.3 F 28.7
3:00 272.4 F 29.3
4:00 272.2 F 28.7

Min 271.0 F 28.0
Max 273.1 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 234.3 F 0.0
10:00 236.1 F 0.0
15:00 228.7 F 0.0
20:00 222.7 F 0.0
25:00 228.9 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 22:36
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:01:36
CYCLE COMPLETE

02/07/2024 01:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____