
Midmark M11 - v1.0.0
Total Cycles: 5083

Name: RITTER 1
Sterilizer ID: V2178198
11/13/2024 10:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.5 F 0.0
2:00 158.6 F 0.2
4:00 171.5 F 0.7
6:00 193.6 F 2.9
8:00 215.0 F 7.0
10:00 237.1 F 11.5
12:00 230.2 F 6.8
14:00 242.8 F 12.2
16:00 239.6 F 10.0
18:00 237.5 F 8.4
20:00 251.6 F 15.5
22:00 263.6 F 22.5
24:00 271.5 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.7
1:00 272.6 F 28.4
2:00 271.4 F 27.6
3:00 272.5 F 28.5
4:00 271.9 F 27.9

Min 271.2 F 27.6
Max 272.6 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.3 F 0.0
5:00 236.5 F 0.0
10:00 215.7 F 0.0
15:00 187.8 F 0.0
20:00 191.3 F 0.0
25:00 189.4 F 0.0
30:00 189.6 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 24:04
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:02:38
CYCLE COMPLETE

11/13/2024 11:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____