

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001658  
11:30 11/12/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:03  
PRESSURIZING 2:10  
STERILIZING 3:14  
136.4 C 221kPa 3:14  
136.3 C 225kPa 3:44  
136.4 C 225kPa 4:14  
136.4 C 226kPa 4:44  
136.4 C 226kPa 5:14  
136.2 C 223kPa 5:44  
136.2 C 224kPa 6:14  
136.4 C 225kPa 6:44  
136.4 C 225kPa 6:44  
VENTING 6:45  
AIR DRYING 7:33  
CYCLE COMPLETE 9:57

-----  
Digital Signature #  
61151062D497C60C  
-----